



**JUF PRODUCTION**

## Medical Face Masks



# Product information

## TECHNICAL DATA SHEET

Date: 12/06/2020

**PRODUCT**

**MEDICAL FACE MASKS**

**CLASSIFICATION**

**TYPE II**

**MODEL**

**ZOGEAR**

**PRODUCT  
DESCRIPTION**

Shingle pleats rectangular face masks with a shapeable nose-piece and two earloops present, one on each side, to hold mask in place.



The product covers the nose, mouth and chin of the mask wearer. Its purpose is to reduce the release of air droplets from the airways of an infected person into the environment to limit the further spread of the infection and to protect the carrier from other environmental hazards.

**MATERIAL**

1st ply	Outer Material	25 g/m2 Non-Woven
2nd ply	Filter Layer	25 g/m2 Melt-Blown filter
3rd ply	Inner Material	25 g/m2 Fluid resistant Non-Woven
Nose-piece	Polypropylene (Non-Metal)	
Earloops	Polyester and Spandex	

Latex free

**QUALITY**

**Norm EN 14683, Type I and Type II limit**

EN 14683 - EFFICACY TESTS	STANDARD / METHOD
Bacterial filtration efficiency (BFE)	EN 14863:2019 Annex B 5.2.2
Breathability	EN 14863:2019 Annex C 5.2.3
Bioburden method validation	ISO 11737-1:2018 Annex B.2
Bioburden (microbiological contamination)	EN 14683:2019 Annex D 5.2.5

**FIXTURE**

3-layer mask body, with nose-piece on the upper edge of the mask and with ultrasonically welded ear loops inside the mask body.

**SIZE**

Mask Body	175 x 95 mm	± 5 mm
Nose-piece	105	± 5 mm
Earloops	165 mm	± 10 mm

**PACKAGING**

PRIMARY PACKAGING	SECONDARY PACKAGING
50 pcs per box Size: 18.5 x 10 x 9cm Weight: 180 g	40x50 pcs per box Size: 52 x 39 x 39.5cm Weight: 7750 g

**SHELF LIFE**

3 years after production date

**STORAGE  
CONDITION**

Store under cool, clean, and dry conditions. Avoid excessive heat (over 40°C or 104°F)

**QUALITY  
ASSURANCE AND  
CONFORMITY**

Directive 93/42/EEC

Regulation (EU) 2017/745


*(Partly. The entry into force of the Regulation postponed to 26 May 2021 - REGULATION (EU) 2020/561)*

EN 14683:2019

**COUNTRY OF  
ORIGIN**

LATVIA













SIA "JUF"  
 Reg. No.: 43603077785  
 Address: Parka street 26-1,  
 Eleja parish, Jelgava county,  
 Latvia, LV-3023

# EC Declaration of Conformity

## EC DECLARATION OF CONFORMITY

No. 20-512-02

	SIA "JUF" declares that this product complies with the following regulations:		
MANUFACTURER	SIA "JUF" Aviacijas street 18, Jelgava, Latvia, LV-3004		
PRODUCT CLASSIFICATION	<b>MEDICAL FACE MASK</b>		
MODEL	<b>ZOGEAR</b>		
PRODUCT DESCRIPTION	The product covers the nose, mouth and chin of the mask wearer. Its purpose is to reduce the release of air droplets from the airways of an infected person into the environment to limit the further spread of the infection and to protect the carrier from other environmental hazards.		
QUALITY ASSURANCE AND CONFORMITY	Directive 93/42/EEC Regulation (EU) 2017/745 <i>(Partly: The entry into force of the Regulation postponed to 26 May 2021 - REGULATION (EU) 2020/561)</i> EN 14683:2019		
	EN 14863:2019	Annex B 5.2.2	Bacterial filtration efficiency (BFE) Test report 20-0647-01
	EN 14863:2019	Annex C 5.2.3	Breathability Test report 20-0647-02
	ISO 11737-1:2018	Annex B.2	Bioburden method validation Test report 20-0647-03
	EN 14683:2019	Annex D 5.2.5	Bioburden (microbiological contamination) Test report 20-0647-04

This declaration of conformity is issued under the sole responsibility of the manufacturer. The object of the declaration identified above is in conformity with the relevant Union harmonisation legislation: Directive 93/42/EEC with its requirements and with the harmonised standards:



We hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced standards and meets all essential requirements of the specified directives

Signed: Baiba Podane  
 Position: Member of the Board  
 Location: Jelgava, Latvia

Signature: \_\_\_\_\_  
 Date of issue: 2020/06/12

# Test Reports

Messrs.  
SIA "JUF"  
Parka street 26-1, Eleja parish.  
LV-3023 JELGAVA COUNTY LATVIA

Zola Predosa, 10/06/2020

Ref. Your Order del 2020

Test Report N°20-0647-01

## DETERMINATION OF BACTERIAL FILTRATION EFFICIENCY (BFE)

### Sample description

# Denomination: Face mask  
# Code: /  
# Lot: /  
# Sterilization: No  
Receipt number: 16211  
Receipt date: 28/05/2020  
Sampling carried out by: SIA "JUF"

### Further information about the sample

Number of tested samples: 5  
Size of the area of the specimens: 50 cm<sup>2</sup>  
Side of the test sample facing the challenge aerosol: the internal part

### Test date

The test was started on 08-06-2020 and was completed on 09-06-2020

### Test method

EN 14683:2019 Annex B

### Equipments and reagents

Vacuum pump "GEO Air Plus"  
Modified Andersen Cascade Impactor "TE-20-830"  
MMAD nebulizer 3,0 ± 0,3 µm  
Culture plates containing TSA

### Summary of method

A negative control is performed by passing air, without addition of the bacterial challenge, through the cascade impactor for 2 minutes.

Then the bacterial challenge of *Staphylococcus Aureus* ATCC 6538, with a concentration of 1,7 x 10<sup>3</sup> to 3,2 x 10<sup>3</sup> UFC/ml, is delivered to the aerosol chamber.

A first positive control is performed, by passing the bacterial challenge through the cascade impactor at a flow rate of 28,3 ± 0,5 l/min for 1 minute. The airflow is maintained through the cascade impactor for 1 additional minute, for a total test time of 2 minutes.

The control plates are removed from the cascade impactor and fresh plates are placed in order to perform the test on the test samples.

Mod. BFE Rv00

Test Report N°20-0647-01

Page 1/2

The specimen is clamped in place between the first stage of the cascade impactor and the inlet cone of the nebulization collector and the procedure used for the positive control is repeated for each of the 5 specimens to be tested.

After the last specimen has been tested, a further positive control run is performed.

Then all the plates are incubated at 37 ± 2°C for a length of time between 24 and 72 hours.

After the incubation, for each specimen and control run, the number of colonies is counted in order to give the total number of CFU collected by the cascade impactor.

The Bacterial Filtration Efficiency (BFE) is calculated for each test specimen, as a percentage, using the following formula:

$$BFE = [(C - T) / C] \times 100$$

where

C is the mean of the total plate counts for the two positive control runs;  
T is the total plate count for the test specimen

### Results

Determination	Collected CFU	BFE (%)	BFE (%) Type I limit	Compliance to Type I limit	BFE (%) Type II and IIR limit	Compliance to Type II and IIR limit
Negative control	0					
Positive control run 1	2858					
Positive control run 2	2872					
Positive control average	2865					
Test 1	5	99,8	≥ 95	In compliance	≥ 98	In compliance
Test 2	7	99,8	≥ 95	In compliance	≥ 98	In compliance
Test 3	9	99,7	≥ 95	In compliance	≥ 98	In compliance
Test 4	11	99,6	≥ 95	In compliance	≥ 98	In compliance
Test 5	8	99,7	≥ 95	In compliance	≥ 98	In compliance
Sample average	8,2	99,7	≥ 95	In compliance	≥ 98	In compliance

The present test report exclusively refers to the referenced test sample.  
If the sample has been sampled by the Customer, the results are referred to the sample as received.  
The present test report may not be partially reproduced without Biochem authorization.

(\*) Data provided by the Customer. The laboratory declines responsibility for such data.

Test verified by: Buriani Giampaolo, PhD.

Issue authorized by:

Head of Laboratory, Giovanni Bassini, Ch. Eng.

END OF TEST REPORT

Mod. BFE Rv00

Test Report N°20-0647-01

Page 2/2

# Test Reports



ANALISE CHIMICO-FISICHE  
MICROBIOLOGICHE  
BIOCOMPATIBILITA'  
CONSULENZA TECNICA  
BIOTECONOLOGIE

Messors  
SIA "JUF"  
Parka street 26-1, Eleje parish,  
LV-3023 JELGAVA COUNTY LATVIA

Zola Predosa, 04/06/2020

Ref. Your Order of 2020

Test Report N°20-0647-02

## DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)

### Sample description

# Denomination: Face mask  
# Code:  
# Lot:  
# Sterilization: No  
Receipt number: 10212  
Receipt date: 26/05/2020  
Sampling carried out by: SIA "JUF"

### Further information about the sample

Number of tested specimens: 5  
Number of tested areas per sample: 5  
General location of the chosen areas to be tested: representative areas are chosen for the test.  
Specimens are taken from one or more samples, depending on the available area.  
Sample preparation: the test is performed on the sample lying flat, wrinkle-free.

Mod. Breathability Fv02

Test Report N°20-0647-02


Page 1/1

VIA BENNE 13 - 40069 ZOLA PREDOSA, BO - TEL. +39-051755295 - FAX +39-051754622  
www.biochem-bcn.com E-mail: info@biochem-bcn.com - C.F. e P. IVA IT 0301810376 - R.E.A. BO-201762



ANALISE CHIMICO-FISICHE  
MICROBIOLOGICHE  
BIOCOMPATIBILITA'  
CONSULENZA TECNICA  
BIOTECONOLOGIE

### Picture of the sample



### Test date

01-06-2020

### Test method

EN 14683 2019 Annex c

### Summary of method

Each specimen is conditioned at  $22 \pm 2^\circ\text{C}$  and  $90 \pm 10\%$  relative humidity for a minimum of 4 hours before the test.  
A device which measures the differential pressure required to draw air through a measured surface area at a constant air flow rate is used to measure the air exchange pressure of the medical face mask material. A digital differential manometer is used to measure the differential pressure. A mass flow meter is used for measurement of the airflow. An electric vacuum pump draws air through the test apparatus and a needle valve is used to adjust the airflow rate.  
Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 l/min.  
The holder is opened and the test specimen is placed across the 25 mm diameter orifice (total area 4,9 cm<sup>2</sup>) between the top and the bottom parts of the holder. Then it is clamped in place using a mechanical clamp with sufficient pressure to avoid air leaks. Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow air. With the specimen in place the flow rate shall be 8 l/min.  
The procedure described is carried out on 5 (or appropriate number) different areas of the mask and the readings averaged.

Mod. Breathability Fv02

Test Report N°20-0647-02

Page 2/2

VIA BENNE 13 - 40069 ZOLA PREDOSA, BO - TEL. +39-051755295 - FAX +39-051754622  
www.biochem-bcn.com E-mail: info@biochem-bcn.com - C.F. e P. IVA IT 0301810376 - R.E.A. BO-201762



ANALISE CHIMICO-FISICHE  
MICROBIOLOGICHE  
BIOCOMPATIBILITA'  
CONSULENZA TECNICA  
BIOTECONOLOGIE

For each test specimen the differential pressure of each tested area is calculated as follows:

$$DP = DP \text{ read} \cdot 4,9$$

when

DP is the Differential Pressure for cm<sup>2</sup> of test material expressed in Pa;  
Dp read is the Differential Pressure for specimen;  
4,9 is the area (in cm<sup>2</sup>) of the test material.

### Results

Determination	DP Read (Pa)	DP (Pa/cm <sup>2</sup> )	DP (Pa/cm <sup>2</sup> ) Type I and II limit	Compliance to Type I and II limit	DP (Pa/cm <sup>2</sup> ) Type III limit	Compliance to Type III limit
Specimen 1 - Pos. A	162	32,1	< 40	Compliant	< 80	Compliant
Specimen 2 - Pos. B	168	34,5	< 40	Compliant	< 80	Compliant
Specimen 3 - Pos. C	156	31,0	< 40	Compliant	< 80	Compliant
Specimen 4 - Pos. D	168	34,5	< 40	Compliant	< 80	Compliant
Specimen 5 - Pos. A	173	35,3	< 40	Compliant	< 80	Compliant
Total average of specimens		33,8	< 40	Compliant	< 80	Compliant

The present test report exclusively refers to the referenced test sample.  
If the sample has been sampled by the Customer, the results are referred to the sample as received.  
The present test report may not be partially reproduced without Biochem authorization.

IRI Data provided by the Customer. The laboratory declines responsibility for such data.

Test verified by: Burtani Giampaolo, PhD

Issue authorized by:  
Head of Laboratory, Giovanni Bassini, Ch. Eng.

END OF TEST REPORT

Mod. Breathability Fv02

Test Report N°20-0647-02

Page 3/3

VIA BENNE 13 - 40069 ZOLA PREDOSA, BO - TEL. +39-051755295 - FAX +39-051754622  
www.biochem-bcn.com E-mail: info@biochem-bcn.com - C.F. e P. IVA IT 0301810376 - R.E.A. BO-201762

Messrs.  
SIA "JUF"  
Parka street 26-1, Eleja parish,  
LV-3023 JELGAVA COUNTY LATVIA

Zola Predosa, 08/06/2020

Ref. Your Order del 2020

# Test Reports

Test Report N° 20-0647-03

DETERMINATION OF A POPULATION OF MICROORGANISMS ON PRODUCTS –  
VALIDATION OF METHOD

## Sample description

# Denomination: Face mask  
# Code: /  
# Lot: /  
# Sterilization: No  
N° of tested samples: 5  
Receipt number: 16213  
Receipt date: 26/05/2020  
Sampling carried out by: SIA "JUF"

The test was started on 03/06/2020 and was completed on 08/06/2020.

## Test method

ISO 11737-1:2018

## Summary of practice

Sterile samples have been contaminated with a know concentration of *Staphylococcus aureus* ATCC 6538 and *Candida albicans* ATCC 10231. Micro-organisms were extracted from samples using sterile physiological saline containing 0,05 % of Tween 80 in mechanical agitation. The extract obtained from each sample was filtrated through a 0,45 µm sterile membrane filter. One half of the filter was incubated on Tryptone Soya Agar (TSA) culture medium for 72 hours at 32°C ± 2°C to evaluate aerobic microorganisms. The other half was incubated on Potato Dextrose Agar (POT) culture medium for 5 days at 22 ± 2°C to evaluate recovered yeasts and moulds . Then was calculated the correction factor for each microorganism.

## Results

*Staphylococcus aureus* ATCC 6538

Sample	Contamination (cfu/sample)	Recovered micro-organism (cfu/sample)	Recovery (%)
1	56	32	57.1
2	56	44	78.6
3	56	22	39.3
4	56	46	82.1
5	56	50	89.3
Mean value		38.8	69.3
Correction factor		1.44	

*Candida albicans* ATCC 10231

Sample	Contamination (cfu/sample)	Recovered micro-organism (cfu/sample)	Recovery (%)
1	294	274	93.2
2	294	204	69.4
3	294	200	68.0
4	294	232	78.9
5	294	170	57.8
Mean value		216	73.5
Correction factor		1.36	

The present test report exclusively refers to the referenced test sample.  
If the sample has been sampled by the Customer, the results are referred to the sample as received.  
The present test report may not be partially reproduced without Biochem authorization.

(#) Data provided by the Customer. The laboratory declines responsibility for such data.

Test verified by: Burianni Giampaolo, PhD.

Issue authorized by:  
Head of Laboratory, Giovanni Bassini, Ch.Eng.

END OF TEST REPORT

# Test Reports

Messrs.  
SIA "JUF"  
Parka street 26-1, Eleja parish.  
LV-3023 JELGAVA COUNTY LATVIA

Zola Predosa, 08/06/2020

Ref. Your Order del 2020

Test Report N° 20-0647-04

## DETERMINATION OF A POPULATION OF MICROORGANISMS ON PRODUCTS

### Sample description

# Denomination: Face mask  
# Code: /  
# Lot: /  
# Sterilization: No  
N° of tested samples: 5  
Receipt number: 16214  
Receipt date: 26/05/2020  
Sampling carried out by: SIA "JUF"

The test was started on 03/06/2020 and was completed on 08/06/2020.

### Test method

ISO 11737-1:2018

### Summary of practice

Samples were aseptically treated. Micro-organisms were extracted from samples using sterile physiological saline containing 0,05 % of Tween 80 in mechanical agitation. The extract was collected and filtered through a 0,45 µm sterile membrane filter. One half of the filter was incubated on Tryptone Soya Agar (TSA) culture medium for 72 hours at 32 ± 2°C in order to evaluate non-selective aerobic bacteria. The other half was incubated on Potato Dextrose Agar (POT) culture medium for 5 days at 22 ± 2°C in order to evaluate yeasts and moulds. Results were multiplied by correction factor (1,44 – 1,36) obtained from the method validation (see test report N° 20-0 647-03).

### Results

Sample	Non-selective aerobic bacteria (cfu/sample)	Moulds (cfu/sample)	Yeast (cfu/sample)
1	10	4	<2
2	44	<2	<2
3	4	<2	<2
4	12	2	<2
5	14	<2	<2
<b>Mean value</b>	<b>16,8</b>	<b>&lt;2,4</b>	<b>&lt;2,0</b>
Correction factor	1,44	1,36	1,36
<b>Corrected value</b>	<b>24,2</b>	<b>&lt;3,3</b>	<b>&lt;2,7</b>
<b>Value per gram</b>	<b>9,5</b>	<b>&lt;2</b>	<b>&lt;2</b>

**SUM OF MICROORGANISMS: <13.5 cfu/gram**

### OPINIONS AND INTERPRETATIONS – Not included in ACCREDIA accreditation

Compliance with EN 14683:2019 5.2.5 Microbial cleanliness (Bioburden): In compliance

The present test report exclusively refers to the referenced test sample. If the sample has been sampled by the Customer, the results are referred to the sample as received. The present test report may not be partially reproduced without Biochem authorization.

(#)Data provided by the Customer. The laboratory declines responsibility for such data.

Test verified by: Buriant Giampaolo, PhD.

Issue authorised by:  
Head of Laboratory Dr. Giovanni Bassini

END OF TEST REPORT

# Packaging



# Packaging

